

Cal. No : XXXXXX

# Calibration Report

Page : 1 / 1

Nomenclature : CARBON DIOXIDE (CO<sub>2</sub>) METER

Cal. Date XX-XX - 2020

Model No. : CENTER 512

Environmental : 22.1 °C

Serial No. : XXXXXXXX

55.5 %RH

## Example

### CO<sub>2</sub>:

Standard	Reading	Deviation
400 ppm	406 ppm	6 ppm
2000 ppm	1963 ppm	-37 ppm

### Temperature :

Standard	Reading	Deviation
22.7 °C	22.7 °C	0.0 °C

### Humidity :

Standard	Reading	Deviation
57.6 %RH	57.1 %RH	-0.5 %RH

## Calibrator

No.	Instrument
1	MolecularAnalysis 7000M
2	rotronic HP22-A
3	AQ COMFORT-2 CO2



# CENTER®

## CENTER TECHNOLOGY CORP.

4/F No.415, Jung-Jeng Road, Shu-Lin Dist.,  
New Taipei City, Taiwan.

TEL : 886 2 2676-3926

E-Mail : center@centertek.com

FAX : 886 2 2676-3925

http : //www.centertek.com

Approved	Tester